

EC Certificate - Full Quality Assurance System

Directive 93/42/EEC on Medical Devices, Annex II excluding Section 4

No.**CE 632392**

Issued To:

**FSSB Chirurgische Nadeln GmbH
Allmendweg 2
79798 Jestetten
Germany**

In respect of:

Design, development and manufacture of sterile PTFE pledgets and sterile, non-absorbable surgical sutures incorporating the following technologies: nylon, polypropylene, polyester, silk and sterile, absorbable surgical sutures incorporating the technology: polyglycolic acid (PGA).

Entwicklung und Herstellung von sterilen Pledgets aus PTFE und von sterilem, nicht resorbierbarem chirurgischem Nahtmaterial aus Nylon, Polypropylen, Polyester, Seide und sterilem, resorbierbarem Nahtmaterial aus Polyglykolsäure (PGA).

on the basis of our examination of the quality assurance system under the requirements of Council Directive 93/42/EEC, Annex II excluding section 4. The quality assurance system meets the requirements of the directive. For the placing on the market of class III products an Annex II section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 0086):



Stewart Brain, Head of Compliance & Risk -
Medical Devices

First Issued: **2016-08-23**Date: **2018-07-10**Expiry Date: **2023-07-06**

...making excellence a habit.™

Page 1 of 1

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body. This approval excludes all products designed and/or manufactured by a third party on behalf of the company named on this certificate, unless specifically agreed with BSI.

This certificate was issued electronically and is bound by the conditions of the contract.